

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072924\
 Data File : VU060248.D
 Acq On : 29 Jul 2024 14:08
 Operator : MD/SY
 Sample : VIBLK123
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK123

Quant Time: Jul 30 04:22:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jul 30 04:21:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	120635	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	128017	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	51232	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	47529	5.663	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	113.200%	
7) Chloroethane-d5	1.911	69	41408	5.310	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	106.200%	
11) 1,1-Dichloroethene-d2	2.560	65	18244	5.567	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	111.400%	
20) 2-Butanone-d5	4.627	46	93908	47.856	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	95.720%	
24) Chloroform-d	5.055	84	82244	4.662	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.200%	
26) 1,2-Dichloroethane-d4	5.695	65	40845	4.973	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.400%	
32) Benzene-d6	5.721	84	162230	4.730	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.600%	
36) 1,2-Dichloropropane-d6	6.685	67	55740	5.100	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.000%	
41) Toluene-d8	7.894	98	129591	4.182	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	8.177	79	15927	4.281	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.600%	
46) 2-Hexanone-d5	8.630	63	64266	38.927	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	77.860%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	44445	4.569	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	44634	4.827	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.600%	
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	11507	1.019	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072924\
Data File : VU060248.D
Acq On : 29 Jul 2024 14:08
Operator : MD/SY
Sample : VIBLK123
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK123

Quant Time: Jul 30 04:22:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jul 30 04:21:24 2024
Response via : Initial Calibration

