

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072924\
 Data File : VU060247.D
 Acq On : 29 Jul 2024 13:44
 Operator : MD/SY
 Sample : P3244-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Jul 30 04:22:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jul 30 04:21:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	133638	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	141614	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	56374	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	50252	5.405	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.000%
7) Chloroethane-d5	1.911	69	42481	4.918	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.400%
11) 1,1-Dichloroethene-d2	2.560	65	19237	5.299	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.000%
20) 2-Butanone-d5	4.627	46	100249	46.116	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.240%
24) Chloroform-d	5.055	84	86261	4.414	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	5.695	65	41558	4.567	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
32) Benzene-d6	5.721	84	171516	4.520	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.400%
36) 1,2-Dichloropropane-d6	6.682	67	57532	4.759	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.894	98	136817	3.991	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	8.177	79	17276	4.198	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.627	63	68236	37.363	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.720%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	44427	4.129	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	46933	4.612	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU072924\
Data File : VU060247.D
Acq On : 29 Jul 2024 13:44
Operator : MD/SY
Sample : P3244-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Jul 30 04:22:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR071824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jul 30 04:21:24 2024
Response via : Initial Calibration

