

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU073018\
 Data File : VU025728.D
 Acq On : 30 Jul 2018 17:25
 Operator : MD/SY
 Sample : J4218-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-15-D

Quant Time: Jul 31 04:28:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 28 02:19:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	175334	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	286155	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	281147	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	170797	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	138828	50.82	ug/l	0.00
Spiked Amount			Recovery	=	101.64%	
35) Dibromofluoromethane	4.89	113	116074	46.64	ug/l	0.00
Spiked Amount			Recovery	=	93.28%	
50) Toluene-d8	7.57	98	411074	47.20	ug/l	0.00
Spiked Amount			Recovery	=	94.40%	
62) 4-Bromofluorobenzene	10.31	95	159133	47.07	ug/l	0.00
Spiked Amount			Recovery	=	94.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	2.83	59	2454	4.10	ug/l #	92
16) Acetone	2.32	43	12883	9.67	ug/l	99
18) Methyl Acetate	2.62	43	3985	1.24	ug/l	93
20) Methylene Chloride	2.70	84	14614	6.19	ug/l	97
25) 2-Butanone	4.27	43	2724	1.50	ug/l #	81
43) Isopropyl Acetate	5.55	43	104054	18.16	ug/l	99
95) Naphthalene	13.74	128	107244	9.50	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU073018\
Data File : VU025728.D
Acq On : 30 Jul 2018 17:25
Operator : MD/SY
Sample : J4218-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-15-D

Quant Time: Jul 31 04:28:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 28 02:19:50 2018
Response via : Initial Calibration

