

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU073119\
 Data File : VU033497.D
 Acq On : 30 Jul 2019 16:54
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: Jul 31 08:18:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 31 08:02:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	349690	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	339829	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	167349	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	117691	46.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.58%
7) Chloroethane-d5	1.67	69	110011	50.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.86%
11) 1,1-Dichloroethene-d2	2.26	63	153847	33.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.68%
21) 2-Butanone-d5	4.16	46	163931	97.35	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.35%
24) Chloroform-d	4.62	84	192647	45.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.29	65	125742	46.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.18%
32) Benzene-d6	5.32	84	401591	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.34%
36) 1,2-Dichloropropane-d6	6.31	67	127477	48.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.54%
41) Toluene-d8	7.55	98	368269	46.65	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.30%
43) trans-1,3-Dichloropropene-	7.83	79	61407	46.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.76%
47) 2-Hexanone-d5	8.30	63	123476	98.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.48%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	185294	45.50	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	11.84	152	146821	47.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.82%

Target Compounds

					Ovalue
14) Carbon disulfide	2.46	76	12508	1.761 ug/L #	94
17) trans-1,2-Dichloroethene	2.97	96	1989	0.838 ug/L	87
42) Toluene	7.62	91	3772	0.358 ug/L	92
46) Tetrachloroethene	8.22	164	1469	0.696 ug/L	79
51) Chlorobenzene	9.10	112	3745	0.543 ug/L	95
52) Ethylbenzene	9.24	91	5188	0.457 ug/L	98
53) m,p-Xylene	9.36	106	2335	0.536 ug/L	72
55) Styrene	9.78	104	3848	0.536 ug/L	98
56) Isopropylbenzene	10.15	105	5203	0.471 ug/L	98
62) 1,3-Dichlorobenzene	11.40	146	7253	1.411 ug/L	97
63) 1,4-Dichlorobenzene	11.49	146	8532	1.598 ug/L	92
65) 1,2-Dichlorobenzene	11.86	146	6758	1.300 ug/L #	94
66) 1,2-Dibromo-3-chloropropan	12.65	75	1283	1.336 ug/L #	82
68) 1,2,4-trichlorobenzene	13.49	180	7267	2.051 ug/L	93
70) 1,2,3-Trichlorobenzene	13.98	180	9607	2.578 ug/L	91

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU073119\
Data File : VU033497.D
Acq On : 30 Jul 2019 16:54
Operator : JC/SP
Sample : VIBLK40
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jul 31 08:18:32 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 31 08:02:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU073119\
Data File : VU033497.D
Acq On : 30 Jul 2019 16:54
Operator : JC/SP
Sample : VIBLK40
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Jul 31 08:18:32 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 31 08:02:52 2019
Response via : Initial Calibration

