

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU073119\
 Data File : VU033533.D
 Acq On : 31 Jul 2019 18:23
 Operator : JC/SP
 Sample : K4103-10DL 40X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB9E6DL

Quant Time: Aug 01 09:05:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 01 07:20:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	290384	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	291546	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	132942	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	100477	47.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.18%
7) Chloroethane-d5	1.67	69	88326	48.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.52%
11) 1,1-Dichloroethene-d2	2.26	63	147845	38.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.16%
21) 2-Butanone-d5	4.15	46	159006	113.71	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.71%
24) Chloroform-d	4.62	84	183425	51.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.66%
26) 1,2-Dichloroethane-d4	5.29	65	122012	54.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.88%
32) Benzene-d6	5.32	84	374662	51.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.68%
36) 1,2-Dichloropropane-d6	6.31	67	122427	54.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.08%
41) Toluene-d8	7.55	98	332907	49.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.32%
43) trans-1,3-Dichloropropene-	7.83	79	55783	49.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.28%
47) 2-Hexanone-d5	8.30	63	114489	106.44	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.44%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	182628	52.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.56%
64) 1,2-Dichlorobenzene-d4	11.85	152	134003	55.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.08%

Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	4.20	96	51827	23.437 ug/L	97
34) Trichloroethene	6.16	95	65521	29.533 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU073119\
Data File : VU033533.D
Acq On : 31 Jul 2019 18:23
Operator : JC/SP
Sample : K4103-10DL 40X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampled :
DB9E6DL

Quant Time: Aug 01 09:05:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 01 07:20:33 2019
Response via : Initial Calibration

