

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU073119\
 Data File : VU033510.D
 Acq On : 30 Jul 2019 23:01
 Operator : JC/SP
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05040

Quant Time: Jul 31 08:12:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 31 08:02:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	49.728	0.5	96	0.00
3 T	Chloromethane	50.000	49.840	0.3	95	0.00
4 S	Vinyl Chloride-d3	50.000	47.006	6.0	94	0.00
5 T	Vinyl chloride	50.000	50.033	-0.1	97	0.00
6 T	Bromomethane	50.000	46.776	6.4	94	0.00
7 S	Chloroethane-d5	50.000	54.614	-9.2	106	0.00
8 T	Chloroethane	50.000	53.764	-7.5	100	0.00
9 T	Trichlorofluoromethane	50.000	50.357	-0.7	97	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	50.000	49.162	1.7	97	0.00
11 S	1,1-Dichloroethene-d2	50.000	47.394	5.2	97	0.00
12 T	1,1-Dichloroethene	50.000	49.299	1.4	97	0.00
13 T	Acetone	100.000	76.554	23.4	71	0.00
14 T	Carbon disulfide	50.000	49.093	1.8	97	0.00
15 T	Methyl Acetate	50.000	50.891	-1.8	97	0.00
16 T	Methylene chloride	50.000	49.618	0.8	96	0.00
17 T	trans-1,2-Dichloroethene	50.000	49.737	0.5	97	0.00
18 T	Methyl tert-butyl Ether	50.000	50.252	-0.5	95	0.00
19 T	1,1-Dichloroethane	50.000	50.402	-0.8	97	0.00
20 T	cis-1,2-Dichloroethene	50.000	50.041	-0.1	97	0.00
21 S	2-Butanone-d5	100.000	100.548	-0.5	98	0.00
22 T	2-Butanone	100.000	100.879	-0.9	90	0.00
23 T	Bromochloromethane	50.000	49.823	0.4	95	0.00
24 S	Chloroform-d	50.000	48.598	2.8	98	0.00
25 T	Chloroform	50.000	50.586	-1.2	96	0.00
26 S	1,2-Dichloroethane-d4	50.000	48.622	2.8	97	0.00
27 T	1,2-Dichloroethane	50.000	50.626	-1.3	96	0.00
28 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
29 T	Cyclohexane	50.000	49.830	0.3	96	0.00
30 T	1,1,1-Trichloroethane	50.000	49.441	1.1	97	0.00
31 T	Carbon tetrachloride	50.000	49.705	0.6	97	0.00
32 S	Benzene-d6	50.000	47.771	4.5	96	0.00
33 T	Benzene	50.000	49.985	0.0	96	0.00
34 T	Trichloroethene	50.000	47.861	4.3	96	0.00
35 T	Methylcyclohexane	50.000	49.341	1.3	95	0.00
36 S	1,2-Dichloropropane-d6	50.000	47.819	4.4	97	0.00
37 T	1,2-Dichloropropane	50.000	50.184	-0.4	98	0.00
38 T	Bromodichloromethane	50.000	49.842	0.3	97	0.00
39 T	cis-1,3-Dichloropropene	50.000	48.014	4.0	93	0.00
40 T	4-Methyl-2-pentanone	100.000	100.316	-0.3	95	0.00
41 S	Toluene-d8	50.000	48.495	3.0	97	0.00
42 T	Toluene	50.000	50.261	-0.5	95	0.00
43 S	trans-1,3-Dichloropropene-d	50.000	46.681	6.6	96	0.00
44 T	trans-1,3-Dichloropropene	50.000	48.116	3.8	92	0.00
45 T	1,1,2-Trichloroethane	50.000	49.327	1.3	96	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU073119\
 Data File : VU033510.D
 Acq On : 30 Jul 2019 23:01
 Operator : JC/SP
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05040

Quant Time: Jul 31 08:12:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 31 08:02:52 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46 T	Tetrachloroethene	50.000	48.134	3.7	94	0.00
47 S	2-Hexanone-d5	100.000	100.919	-0.9	98	0.00
48 T	2-Hexanone	100.000	99.063	0.9	92	0.00
49 T	Dibromochloromethane	50.000	49.338	1.3	95	0.00
50 T	1,2-Dibromoethane	50.000	49.641	0.7	96	0.00
51 T	Chlorobenzene	50.000	49.362	1.3	96	0.00
52 T	Ethylbenzene	50.000	50.432	-0.9	96	0.00
53 T	m,p-Xylene	50.000	49.973	0.1	96	0.00
54 T	o-xylene	50.000	51.237	-2.5	95	0.00
55 T	Styrene	50.000	51.530	-3.1	95	0.00
56 T	Isopropylbenzene	50.000	50.198	-0.4	95	0.00
57 S	1,1,2,2-Tetrachloroethane-d	50.000	47.081	5.8	98	0.00
58 T	1,1,2,2-Tetrachloroethane	50.000	49.506	1.0	96	0.00
59	1,2,3-Trichloropropane	50.000	49.269	1.5	95	0.00
60 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
61 T	Bromoform	50.000	49.007	2.0	95	0.00
62 T	1,3-Dichlorobenzene	50.000	48.527	2.9	96	0.00
63 T	1,4-Dichlorobenzene	50.000	47.397	5.2	95	0.00
64 S	1,2-Dichlorobenzene-d4	50.000	45.981	8.0	98	0.00
65 T	1,2-Dichlorobenzene	50.000	48.623	2.8	95	0.00
66 T	1,2-Dibromo-3-chloropropane	50.000	49.145	1.7	97	0.00
67	1,3,5-Trichlorobenzene	50.000	46.734	6.5	93	0.00
68 T	1,2,4-trichlorobenzene	50.000	48.147	3.7	94	0.00
69	Naphthalene	50.000	48.607	2.8	91	0.00
70 T	1,2,3-Trichlorobenzene	50.000	48.082	3.8	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0