

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU073124\
 Data File : VU060280.D
 Acq On : 31 Jul 2024 13:37
 Operator : MD/SY
 Sample : P3347-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Aug 01 00:21:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 01 00:19:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	229148	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	234962	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	91830	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	93447	63.514	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 127.020%		
7) Chloroethane-d5	1.901	69	70045	56.600	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 113.200%		
11) 1,1-Dichloroethene-d2	2.557	63	107469	41.116	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 82.240%		
21) 2-Butanone-d5	4.611	46	114803	112.278	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 112.280%		
24) Chloroform-d	5.055	84	143389	50.326	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 100.660%		
26) 1,2-Dichloroethane-d4	5.695	65	88635	52.320	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.640%		
32) Benzene-d6	5.721	84	294807	48.998	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 98.000%		
36) 1,2-Dichloropropane-d6	6.682	67	99038	53.185	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 106.360%		
41) Toluene-d8	7.891	98	246227	43.034	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 86.060%		
43) trans-1,3-Dichloroprop...	8.174	79	40650	46.651	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 93.300%		
47) 2-Hexanone-d5	8.627	63	66449	84.671	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 84.670%		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	139959	48.968	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 97.940%		
66) 1,2-Dichlorobenzene-d4	12.187	152	85489	50.661	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 101.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU073124\
Data File : VU060280.D
Acq On : 31 Jul 2024 13:37
Operator : MD/SY
Sample : P3347-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Aug 01 00:21:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 01 00:19:02 2024
Response via : Initial Calibration

