

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080118\
 Data File : VU025739.D
 Acq On : 31 Jul 2018 15:13
 Operator : MD/SY
 Sample : VU0801MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0801MBL01

Quant Time: Jul 31 17:17:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 13:11:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	187024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	291548	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	283497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	164261	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	141299	48.49	ug/l	0.00
Spiked Amount						
			Recovery	=	96.98%	
35) Dibromofluoromethane	4.89	113	123378	48.66	ug/l	0.00
Spiked Amount						
			Recovery	=	97.32%	
50) Toluene-d8	7.57	98	424613	47.85	ug/l	0.00
Spiked Amount						
			Recovery	=	95.70%	
62) 4-Bromofluorobenzene	10.31	95	162303	47.12	ug/l	0.00
Spiked Amount						
			Recovery	=	94.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080118\
Data File : VU025739.D
Acq On : 31 Jul 2018 15:13
Operator : MD/SY
Sample : VU0801MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0801MBL01

Quant Time: Jul 31 17:17:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 13:11:52 2018
Response via : Initial Calibration

