

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080118\
 Data File : VU025775.D
 Acq On : 01 Aug 2018 14:27
 Operator : MD/SY
 Sample : J4262-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB-073118

Quant Time: Aug 02 04:33:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 31 16:07:47 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	177221	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	287047	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	283809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	169464	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	141379	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
35) Dibromofluoromethane	4.89	113	118941	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
50) Toluene-d8	7.57	98	416679	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
62) 4-Bromofluorobenzene	10.31	95	160219	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
Target Compounds						
16) Acetone	2.32	43	3766	2.797	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080118\
Data File : VU025775.D
Acq On : 01 Aug 2018 14:27
Operator : MD/SY
Sample : J4262-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB-073118

Quant Time: Aug 02 04:33:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 31 16:07:47 2018
Response via : Initial Calibration

