

Data File : VU025782.D
Acq On : 01 Aug 2018 17:29
Operator : MD/SY
Sample : J4262-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-073118

Manual Integrations
APPROVED

MMDadoda
8/2/2018 9:45:07 AM

Quant Time: Aug 11 03:42:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 31 16:07:47 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	317247	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	312633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	198503	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.31	65	154829	49.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.22%	
35) Dibromofluoromethane	4.89	113	131765	47.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.50%	
50) Toluene-d8	7.57	98	462038	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	
62) 4-Bromofluorobenzene	10.31	95	187695	50.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.16%	

Target Compounds

						Qvalue
16) Acetone	2.33	43	7122	4.634	ug/l	96
19) Methyl tert-butyl Ether	3.00	73	5866	0.745	ug/l	99
40) Benzene	5.39	78	136249	12.572	ug/l	99
52) Toluene	7.64	92	6101	0.906	ug/l	100
67) Ethyl Benzene	9.25	91	49021	3.672	ug/l	98
68) m/p-Xylenes	9.38	106	17575	3.429	ug/l	93
69) o-Xylene	9.78	106	11650	2.325	ug/l	97
73) Isopropylbenzene	10.17	105	98662	7.282	ug/l	98
78) n-propylbenzene	10.59	91	32437	2.051	ug/l	99
80) 1,3,5-Trimethylbenzene	10.78	105	14678	1.276	ug/l	96
84) 1,2,4-Trimethylbenzene	11.15	105	31675	2.717	ug/l	98
86) p-Isopropyltoluene	11.48	119	4651	0.372	ug/l #	72
89) n-Butylbenzene	11.89	91	9354m	0.811	ug/l	
95) Naphthalene	13.75	128	4477789	297.444	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU025782.D
Acq On : 01 Aug 2018 17:29
Operator : MD/SY
Sample : J4262-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DUP-073118

Manual Integrations
APPROVED

MMDadoda
8/2/2018 9:45:07 AM

Quant Time: Aug 11 03:42:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 31 16:07:47 2018
Response via : Initial Calibration

