

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU080118\
 Data File : VU025758.D
 Acq On : 31 Jul 2018 22:48
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 01 06:05:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 31 16:07:47 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	0.756	0.791	-4.6	101	0.00
3 P	Chloromethane	0.722	0.647	10.4	94	0.00
4 C	Vinyl Chloride	0.680	0.670	1.5#	104	0.00
5 T	Bromomethane	0.332	0.348	-4.8	107	0.00
6 T	Chloroethane	0.397	0.383	3.5	104	0.00
7 T	Trichlorofluoromethane	0.983	0.986	-0.3	104	0.00
8 T	Diethyl Ether	0.356	0.347	2.5	105	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.569	0.575	-1.1	105	0.00
10 T	Methyl Iodide	0.440	0.515	-17.0	120	0.00
11 T	Tert butyl alcohol	0.171	0.154	9.9	98	0.00
12 CM	1,1-Dichloroethene	0.537	0.542	-0.9#	109	0.00
13 T	Acrolein	0.068	0.077	-13.2	136	0.00
14 T	Allyl chloride	0.965	0.905	6.2	106	0.00
15 T	Acrylonitrile	0.341	0.340	0.3	103	0.00
16 T	Acetone	0.380	0.304	20.0#	85	0.00
17 T	Carbon Disulfide	1.737	1.642	5.5	102	0.00
18 T	Methyl Acetate	0.914	1.047	-14.6	113	0.00
19 T	Methyl tert-butyl Ether	1.945	1.996	-2.6	107	0.00
20 T	Methylene Chloride	0.673	0.637	5.3	104	0.00
21 T	trans-1,2-Dichloroethene	0.621	0.609	1.9	105	0.00
22 T	Diisopropyl ether	1.901	1.951	-2.6	107	0.00
23 T	Vinyl Acetate	1.657	1.632	1.5	103	0.00
24 P	1,1-Dichloroethane	1.141	1.152	-1.0	106	0.00
25 T	2-Butanone	0.517	0.482	6.8	97	0.00
26 T	2,2-Dichloropropane	1.037	0.874	15.7	90	0.00
27 T	cis-1,2-Dichloroethene	0.700	0.687	1.9	105	0.00
28 T	Bromochloromethane	0.510	0.515	-1.0	103	0.00
29 T	Tetrahydrofuran	0.310	0.303	2.3	102	0.00
30 C	Chloroform	1.205	1.201	0.3#	106	0.00
31 T	Cyclohexane	1.194	0.983	17.7	103	0.00
32 T	1,1,1-Trichloroethane	1.085	1.097	-1.1	106	0.00
33 S	1,2-Dichloroethane-d4	0.779	0.774	0.6	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
35 S	Dibromofluoromethane	0.435	0.426	2.1	106	0.00
36 T	1,1-Dichloropropene	0.584	0.589	-0.9	106	0.00
37 T	Ethyl Acetate	0.606	0.596	1.7	102	0.00
38 T	Carbon Tetrachloride	0.662	0.662	0.0	106	0.00
39 T	Methylcyclohexane	0.670	0.655	2.2	103	0.00
40 TM	Benzene	1.708	1.691	1.0	107	0.00
41 T	Methacrylonitrile	0.337	0.339	-0.6	105	0.00
42 TM	1,2-Dichloroethane	0.658	0.649	1.4	108	0.00
43 T	Isopropyl Acetate	1.001	0.988	1.3	105	0.00
44 TM	Trichloroethene	0.476	0.477	-0.2	109	0.00
45 C	1,2-Dichloropropane	0.445	0.439	1.3#	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU080118\
 Data File : VU025758.D
 Acq On : 31 Jul 2018 22:48
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 01 06:05:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 31 16:07:47 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.323	0.319	1.2	108	0.00
47 T	Bromodichloromethane	0.616	0.617	-0.2	106	0.00
48 T	Methyl methacrylate	0.491	0.507	-3.3	106	0.00
49 T	1,4-Dioxane	0.012	0.011	8.3	97	0.00
50 S	Toluene-d8	1.522	1.530	-0.5	105	0.00
51 T	4-Methyl-2-Pentanone	0.627	0.629	-0.3	104	0.00
52 CM	Toluene	1.061	1.082	-2.0#	107	0.00
53 T	t-1,3-Dichloropropene	0.691	0.692	-0.1	105	0.00
54 T	cis-1,3-Dichloropropene	0.710	0.721	-1.5	106	0.00
55 T	1,1,2-Trichloroethane	0.441	0.444	-0.7	107	0.00
56 T	Ethyl methacrylate	0.630	0.655	-4.0	108	0.00
57 T	1,3-Dichloropropane	0.742	0.740	0.3	108	0.00
58 T	2-Chloroethyl Vinyl ether	0.260	0.260	0.0	101	0.00
59 T	2-Hexanone	0.491	0.486	1.0	102	0.00
60 T	Dibromochloromethane	0.520	0.519	0.2	107	0.00
61 T	1,2-Dibromoethane	0.490	0.491	-0.2	106	0.00
62 S	4-Bromofluorobenzene	0.591	0.605	-2.4	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.440	0.445	-1.1	108	0.00
65 PM	Chlorobenzene	1.289	1.286	0.2	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.466	0.480	-3.0	107	0.00
67 C	Ethyl Benzene	2.135	2.174	-1.8#	108	0.00
68 T	m/p-Xylenes	0.820	0.859	-4.8	107	0.00
69 T	o-Xylene	0.801	0.819	-2.2	106	0.00
70 T	Styrene	1.294	1.369	-5.8	107	0.00
71 P	Bromoform	0.404	0.424	-5.0	109	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.413	3.511	-2.9	107	0.00
74 T	N-amyl acetate	1.393	1.433	-2.9	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.180	1.158	1.9	105	0.00
76 T	1,2,3-Trichloropropane	1.102	1.058	4.0	102	0.00
77 T	Bromobenzene	0.941	0.953	-1.3	109	0.00
78 T	n-propylbenzene	3.983	4.126	-3.6	107	0.00
79 T	2-Chlorotoluene	2.333	2.410	-3.3	108	0.00
80 T	1,3,5-Trimethylbenzene	2.898	3.032	-4.6	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.439	0.368	16.2	101	0.02
82 T	4-Chlorotoluene	2.750	2.866	-4.2	107	0.00
83 T	tert-Butylbenzene	2.880	2.941	-2.1	107	0.00
84 T	1,2,4-Trimethylbenzene	2.937	3.078	-4.8	106	0.00
85 T	sec-Butylbenzene	3.516	3.604	-2.5	105	0.00
86 T	p-Isopropyltoluene	3.148	3.261	-3.6	105	0.00
87 T	1,3-Dichlorobenzene	1.747	1.766	-1.1	108	0.00
88 T	1,4-Dichlorobenzene	1.821	1.790	1.7	107	0.00
89 T	n-Butylbenzene	2.905	2.911	-0.2	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU080118\
Data File : VU025758.D
Acq On : 31 Jul 2018 22:48
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Aug 01 06:05:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 31 16:07:47 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.560	0.543	3.0	104	0.00
91 T	1,2-Dichlorobenzene	1.713	1.747	-2.0	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.283	-1.1	106	0.00
93 T	1,2,4-Trichlorobenzene	1.156	1.222	-5.7	107	0.00
94 T	Hexachlorobutadiene	0.609	0.552	9.4	98	0.00
95 T	Naphthalene	3.272	3.729	-14.0	108	0.00
96 T	1,2,3-Trichlorobenzene	1.153	1.221	-5.9	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6