


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU080118\
Data File : VU025736.D
Acq On : 31 Jul 2018 12:44
Operator : MD/SY
Sample : BFB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Title : SW846 8260
Last Update : Tue Jul 31 16:07:47 2018



AutoFind: Scans 3022, 3023, 3024; Background Corrected with Scan 3010

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	7594	PASS
75	95	30	60	54.3	22253	PASS
95	95	100	100	100.0	41013	PASS
96	95	5	9	7.4	3017	PASS
173	174	0.00	2	1.2	446	PASS
174	95	50	100	89.4	36666	PASS
175	174	5	9	8.1	2966	PASS
176	174	95	101	95.1	34861	PASS
177	176	5	9	6.0	2093	PASS