

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080119\
 Data File : VU033550.D
 Acq On : 01 Aug 2019 15:35
 Operator : JC/SP
 Sample : K4116-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 JLLH2

Quant Time: Aug 02 02:57:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 02 02:28:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	310547	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	307056	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	141994	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	142502	63.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	126.22%
7) Chloroethane-d5	1.67	69	132982	68.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	137.30%#
11) 1,1-Dichloroethene-d2	2.26	63	193792	47.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.58%
21) 2-Butanone-d5	4.16	46	190016	127.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	127.06%
24) Chloroform-d	4.62	84	216921	57.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.64%
26) 1,2-Dichloroethane-d4	5.29	65	141498	59.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	118.08%
32) Benzene-d6	5.32	84	449434	59.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	118.10%
36) 1,2-Dichloropropane-d6	6.31	67	141875	59.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.92%
41) Toluene-d8	7.55	98	400498	56.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.30%
43) trans-1,3-Dichloropropene-	7.83	79	64654	54.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.24%
47) 2-Hexanone-d5	8.30	63	131334	115.93	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	115.93%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	205352	55.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.62%
64) 1,2-Dichlorobenzene-d4	11.85	152	154199	59.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080119\
Data File : VU033550.D
Acq On : 01 Aug 2019 15:35
Operator : JC/SP
Sample : K4116-20
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
JLLH2

Quant Time: Aug 02 02:57:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM073119WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 02 02:28:15 2019
Response via : Initial Calibration

