

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080124\
 Data File : VU060287.D
 Acq On : 01 Aug 2024 10:05
 Operator : MD/SY
 Sample : VU0801MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK103

Quant Time: Aug 02 02:32:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 02 02:32:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	219486	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	220691	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	86168	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	85812	60.892	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	121.780%
7) Chloroethane-d5	1.904	69	66091	55.756	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.520%
11) 1,1-Dichloroethene-d2	2.560	63	99999	39.942	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.880%
21) 2-Butanone-d5	4.615	46	105944	108.175	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.180%
24) Chloroform-d	5.052	84	136953	50.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.360%
26) 1,2-Dichloroethane-d4	5.695	65	82793	51.023	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.040%
32) Benzene-d6	5.718	84	276450	48.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.840%
36) 1,2-Dichloropropane-d6	6.682	67	94031	53.761	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.520%
41) Toluene-d8	7.891	98	229932	42.784	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.560%
43) trans-1,3-Dichloroprop...	8.174	79	37786	46.168	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.340%
47) 2-Hexanone-d5	8.627	63	59130	80.218	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.220%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	136094	50.695	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.380%
66) 1,2-Dichlorobenzene-d4	12.187	152	78545	49.605	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080124\
Data File : VU060287.D
Acq On : 01 Aug 2024 10:05
Operator : MD/SY
Sample : VU0801MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK103

Quant Time: Aug 02 02:32:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM070224WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 02 02:32:23 2024
Response via : Initial Calibration

