

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080218\
 Data File : VU025789.D
 Acq On : 02 Aug 2018 10:22
 Operator : MD/SY
 Sample : VU0802WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK56

Quant Time: Aug 03 04:29:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 03 04:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	297884	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	277952	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	146420	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	86965	51.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.28%
7) Chloroethane-d5	1.68	69	77607	55.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	110.64%
11) 1,1-Dichloroethene-d2	2.27	63	141239	39.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.78%
21) 2-Butanone-d5	4.18	46	175302	133.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	133.52%#
24) Chloroform-d	4.65	84	218051	57.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.12%
26) 1,2-Dichloroethane-d4	5.31	65	152126	60.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.46%
32) Benzene-d6	5.35	84	412700	57.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.78%
36) 1,2-Dichloropropane-d6	6.33	67	136766	59.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	119.08%
41) Toluene-d8	7.57	98	389783	57.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.04%
43) trans-1,3-Dichloropropene-	7.85	79	68246	58.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	116.34%
47) 2-Hexanone-d5	8.31	63	124417	129.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	129.36%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	193080	60.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	120.20%#
64) 1,2-Dichlorobenzene-d4	11.86	152	171221	57.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080218\
Data File : VU025789.D
Acq On : 02 Aug 2018 10:22
Operator : MD/SY
Sample : VU0802WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK56

Quant Time: Aug 03 04:29:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 03 04:24:02 2018
Response via : Initial Calibration

