

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080218\
 Data File : VU025803.D
 Acq On : 02 Aug 2018 19:38
 Operator : MD/SY
 Sample : J4265-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AB2

Quant Time: Aug 03 06:08:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 03 05:43:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	312228	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	295078	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	148983	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	79030	44.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.68%
7) Chloroethane-d5	1.68	69	64279	43.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.44%
11) 1,1-Dichloroethene-d2	2.27	63	120698	32.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.06%
21) 2-Butanone-d5	4.18	46	141462	102.80	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.80%
24) Chloroform-d	4.65	84	171387	43.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.34%
26) 1,2-Dichloroethane-d4	5.31	65	120628	45.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.88%
32) Benzene-d6	5.35	84	336122	44.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.82%
36) 1,2-Dichloropropane-d6	6.34	67	109831	45.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.08%
41) Toluene-d8	7.57	98	321520	44.30	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.60%
43) trans-1,3-Dichloropropene-	7.85	79	53031	42.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.16%
47) 2-Hexanone-d5	8.31	63	100927	98.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.85%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	148003	43.40	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	132230	43.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.58%

Target Compounds					Ovalue
13) Acetone	2.33	43	5982	3.70 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080218\
Data File : VU025803.D
Acq On : 02 Aug 2018 19:38
Operator : MD/SY
Sample : J4265-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AB2

Quant Time: Aug 03 06:08:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 03 05:43:06 2018
Response via : Initial Calibration

