

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080218\
 Data File : VU025808.D
 Acq On : 02 Aug 2018 21:35
 Operator : MD/SY
 Sample : J4265-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AB7

Quant Time: Aug 03 06:37:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 03 05:43:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	300889	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	285537	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	143007	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	76790	44.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.40%
7) Chloroethane-d5	1.68	69	61796	43.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.22%
11) 1,1-Dichloroethene-d2	2.27	63	124152	34.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.44%
21) 2-Butanone-d5	4.18	46	131237	98.96	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.96%
24) Chloroform-d	4.65	84	165656	43.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.58%
26) 1,2-Dichloroethane-d4	5.31	65	114878	45.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.80%
32) Benzene-d6	5.35	84	326322	44.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.12%
36) 1,2-Dichloropropane-d6	6.34	67	104577	44.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.64%
41) Toluene-d8	7.57	98	307537	43.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.58%
43) trans-1,3-Dichloropropene-	7.85	79	51725	42.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.82%
47) 2-Hexanone-d5	8.31	63	94622	95.77	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.77%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	142932	43.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.62%
64) 1,2-Dichlorobenzene-d4	11.86	152	128357	44.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.58%

Target Compounds					Ovalue
13) Acetone	2.32	43	4537	2.91 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080218\
Data File : VU025808.D
Acq On : 02 Aug 2018 21:35
Operator : MD/SY
Sample : J4265-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AB7

Quant Time: Aug 03 06:37:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 03 05:43:06 2018
Response via : Initial Calibration

