

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080318\
 Data File : VU025813.D
 Acq On : 03 Aug 2018 12:14
 Operator : MD/SY
 Sample : VU0803WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0803WBL01

Quant Time: Aug 04 05:37:54 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072518DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 27 15:10:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.74	96	31482	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	10195	0.87	ug/l	0.00
Spiked Amount 1.000			Recovery	=	87.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	11114	0.86	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.00%	

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080318\
Data File : VU025813.D
Acq On : 03 Aug 2018 12:14
Operator : MD/SY
Sample : VU0803WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0803WBL01

Quant Time: Aug 04 05:37:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072518DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jul 27 15:10:58 2018
Response via : Initial Calibration

