

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080318\
 Data File : VU025815.D
 Acq On : 03 Aug 2018 13:14
 Operator : MD/SY
 Sample : J4229-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BLUEBERRY-LAKE-PW

Quant Time: Aug 04 05:41:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U072518DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Jul 27 15:10:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.75	96	32735	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	11105	0.91	ug/l	0.00
Spiked Amount 1.000			Recovery	=	91.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	12018	0.89	ug/l	0.00
Spiked Amount 1.000			Recovery	=	89.00%	

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080318\
Data File : VU025815.D
Acq On : 03 Aug 2018 13:14
Operator : MD/SY
Sample : J4229-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BLUEBERRY-LAKE-PW

Quant Time: Aug 04 05:41:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U072518DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Jul 27 15:10:58 2018
Response via : Initial Calibration

