

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080322\
 Data File : VU050053.D
 Acq On : 03 Aug 2022 14:59
 Operator : SY/MD
 Sample : N4027-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJH5

Quant Time: Aug 04 01:45:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Aug 04 01:44:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	81517	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	81673	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	31082	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	31727	4.533	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.600%
7) Chloroethane-d5	1.909	69	26042	5.180	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.600%
11) 1,1-Dichloroethene-d2	2.565	65	11631	4.822	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.400%
20) 2-Butanone-d5	4.632	46	80588	51.724	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.440%
24) Chloroform-d	5.063	84	58801	5.077	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
26) 1,2-Dichloroethane-d4	5.703	65	30350	5.200	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.000%
32) Benzene-d6	5.729	84	114411	4.741	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
36) 1,2-Dichloropropane-d6	6.693	67	39065	4.941	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.800%
41) Toluene-d8	7.899	98	91581	4.269	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.400%
43) trans-1,3-Dichloroprop...	8.182	79	12535	4.524	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%
46) 2-Hexanone-d5	8.635	63	43052	41.502	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.000%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	33095	4.982	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	30226	5.243	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080322\
Data File : VU050053.D
Acq On : 03 Aug 2022 14:59
Operator : SY/MD
Sample : N4027-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJH5

Quant Time: Aug 04 01:45:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Aug 04 01:44:33 2022
Response via : Initial Calibration

