

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080418\
 Data File : VU025835.D
 Acq On : 03 Aug 2018 21:54
 Operator : MD/SY
 Sample : J4340-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E5F75

Quant Time: Aug 04 00:57:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 03 23:51:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	312804	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	292527	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	132273	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	80560	45.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.22%
7) Chloroethane-d5	1.68	69	68675	46.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.24%
11) 1,1-Dichloroethene-d2	2.27	63	129793	34.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.82%
21) 2-Butanone-d5	4.18	46	142083	103.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.06%
24) Chloroform-d	4.65	84	183400	46.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.22%
26) 1,2-Dichloroethane-d4	5.31	65	126277	48.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.00%
32) Benzene-d6	5.35	84	352541	46.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.98%
36) 1,2-Dichloropropane-d6	6.34	67	115483	47.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.54%
41) Toluene-d8	7.57	98	317345	44.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.22%
43) trans-1,3-Dichloropropene-	7.85	79	55519	44.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.92%
47) 2-Hexanone-d5	8.31	63	102222	100.99	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.99%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	153434	45.38	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.76%
64) 1,2-Dichlorobenzene-d4	11.86	152	123403	46.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.06%

Target Compounds

					Ovalue
13) Acetone	2.32	43	5758	3.56 ug/L	94
20) cis-1,2-Dichloroethene	4.23	96	2765	1.21 ug/L	95
34) Trichloroethene	6.19	95	3421	1.53 ug/L	81
46) Tetrachloroethene	8.23	164	36348	18.80 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080418\
Data File : VU025835.D
Acq On : 03 Aug 2018 21:54
Operator : MD/SY
Sample : J4340-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E5F75

Quant Time: Aug 04 00:57:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 03 23:51:54 2018
Response via : Initial Calibration

