

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080418\
 Data File : VU025851.D
 Acq On : 04 Aug 2018 04:07
 Operator : MD/SY
 Sample : J4261-09ME
 Misc : 3.54g/5.0mL/100uL/5.00mL//MSVOA_U/MEOH
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E5F96ME

Manual Integrations
 APPROVED

sam
 8/6/2018 12:30:51 PM

Quant Time: Aug 04 05:27:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Aug 04 04:39:04 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	272646	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	259040	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	137762	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	39029	25.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	50.14%#
7) Chloroethane-d5	1.64	69	22542m	17.56	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	35.12%#
11) 1,1-Dichloroethene-d2	2.24	63	61297	18.92	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	37.84%#
21) 2-Butanone-d5	4.20	46	68322	56.86	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	56.86%
24) Chloroform-d	4.65	84	86310	24.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	49.78%#
26) 1,2-Dichloroethane-d4	5.31	65	63979	27.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	55.80%#
32) Benzene-d6	5.34	84	172199	25.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	51.84%#
36) 1,2-Dichloropropane-d6	6.33	67	55273	25.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	51.64%#
41) Toluene-d8	7.57	98	157112	24.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	49.32%#
43) trans-1,3-Dichloropropene-	7.85	79	25950	23.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	47.46%#
47) 2-Hexanone-d5	8.33	63	47116	52.56	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	52.56%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	81122	27.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	54.20%#
64) 1,2-Dichlorobenzene-d4	11.87	152	70886	25.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	50.78%#

Target Compounds					Ovalue
46) Tetrachloroethene	8.23	164	9187	5.37 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080418\
Data File : VU025851.D
Acq On : 04 Aug 2018 04:07
Operator : MD/SY
Sample : J4261-09ME
Misc : 3.54g/5.0mL/100uL/5.00mL//MSVOA_U/MEOH
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E5F96ME

Manual Integrations
APPROVED

sam
8/6/2018 12:30:51 PM

Quant Time: Aug 04 05:27:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Aug 04 04:39:04 2018
Response via : Initial Calibration

