

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080422\
 Data File : VU050073.D
 Acq On : 04 Aug 2022 20:09
 Operator : SY/MD
 Sample : N4022-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5D80

Quant Time: Aug 05 04:55:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 05 04:52:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	81483	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	83341	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	33223	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	33420	4.928	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.600%
7) Chloroethane-d5	1.912	69	28492	5.208	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.200%
11) 1,1-Dichloroethene-d2	2.565	65	12553	4.737	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.800%
20) 2-Butanone-d5	4.636	46	100883	61.007	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.020%
24) Chloroform-d	5.063	84	67382	5.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.800%
26) 1,2-Dichloroethane-d4	5.703	65	34916	5.392	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	5.729	84	132704	5.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
36) 1,2-Dichloropropane-d6	6.690	67	46397	5.417	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.400%
41) Toluene-d8	7.899	98	109708	4.776	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
43) trans-1,3-Dichloroprop...	8.182	79	15135	4.932	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.600%
46) 2-Hexanone-d5	8.635	63	57360	52.704	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.400%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	39797	5.291	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	12.195	152	37526	5.765	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.200%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.604	62	104582	11.947	ug/L	99
13) Acetone	2.652	43	4152	4.509	ug/L	95
18) trans-1,2-Dichloroethene	3.353	96	3151	0.500	ug/L	97
22) cis-1,2-Dichloroethene	4.668	96	21847	3.121	ug/L	99
42) Toluene	7.970	91	5944	0.203	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080422\
Data File : VU050073.D
Acq On : 04 Aug 2022 20:09
Operator : SY/MD
Sample : N4022-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5D80

Quant Time: Aug 05 04:55:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 05 04:52:21 2022
Response via : Initial Calibration

