

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080422\
 Data File : VU050077.D
 Acq On : 04 Aug 2022 21:47
 Operator : SY/MD
 Sample : N4022-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F5E55

Quant Time: Aug 05 04:57:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Aug 05 04:52:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	82976	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	85654	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	38404	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	32083	4.646	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.000%
7) Chloroethane-d5	1.912	69	26919	4.832	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.565	65	12983	4.811	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.200%
20) 2-Butanone-d5	4.636	46	94568	56.159	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.320%
24) Chloroform-d	5.063	84	64914	5.054	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
26) 1,2-Dichloroethane-d4	5.703	65	33967	5.151	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.729	84	129562	4.847	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	6.690	67	43988	4.997	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.000%
41) Toluene-d8	7.899	98	106488	4.511	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
43) trans-1,3-Dichloroprop...	8.179	79	14595	4.627	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.600%
46) 2-Hexanone-d5	8.635	63	56373	50.399	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.800%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	37584	4.862	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.192	152	40252	5.349	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.000%
Target Compounds						Qvalue
16) Methylene chloride	3.047	84	15430	1.761	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080422\
Data File : VU050077.D
Acq On : 04 Aug 2022 21:47
Operator : SY/MD
Sample : N4022-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F5E55

Quant Time: Aug 05 04:57:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Aug 05 04:52:21 2022
Response via : Initial Calibration

