

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU080518\
 Data File : VU025864.D
 Acq On : 04 Aug 2018 11:29
 Operator : MD/SY
 Sample : VU0805MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK42

Quant Time: Aug 06 01:30:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Aug 06 01:30:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	284521	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	271739	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	132231	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	66952	41.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.44%
7) Chloroethane-d5	1.68	69	58416	43.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.20%
11) 1,1-Dichloroethene-d2	2.27	63	108847	32.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.38%
21) 2-Butanone-d5	4.18	46	126584	100.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.94%
24) Chloroform-d	4.65	84	169015	46.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.42%
26) 1,2-Dichloroethane-d4	5.31	65	117431	49.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.16%
32) Benzene-d6	5.35	84	321472	46.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.26%
36) 1,2-Dichloropropane-d6	6.33	67	104588	46.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.14%
41) Toluene-d8	7.57	98	296470	44.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.72%
43) trans-1,3-Dichloropropene-	7.85	79	52882	46.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.20%
47) 2-Hexanone-d5	8.31	63	86843	92.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.36%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	143502	45.69	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.38%
64) 1,2-Dichlorobenzene-d4	11.86	152	126929	47.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU080518\
Data File : VU025864.D
Acq On : 04 Aug 2018 11:29
Operator : MD/SY
Sample : VU0805MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK42

Quant Time: Aug 06 01:30:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM072018WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Aug 06 01:30:36 2018
Response via : Initial Calibration

