

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080522\
 Data File : VU050101.D
 Acq On : 05 Aug 2022 17:05
 Operator : SY/MD
 Sample : N4027-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJH6

Quant Time: Aug 06 04:36:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 06 04:32:47 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	77968	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	80133	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	29432	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	28529	4.397	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.000%	
7) Chloroethane-d5	1.909	69	24831	4.743	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.800%	
11) 1,1-Dichloroethene-d2	2.565	65	11438	4.510	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.200%	
20) 2-Butanone-d5	4.623	46	90053	56.913	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	113.820%	
24) Chloroform-d	5.067	84	58051	4.810	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
26) 1,2-Dichloroethane-d4	5.703	65	32984	5.323	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.400%	
32) Benzene-d6	5.729	84	116613	4.663	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.200%	
36) 1,2-Dichloropropane-d6	6.694	67	40800	4.955	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.000%	
41) Toluene-d8	7.899	98	95616	4.329	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.600%	
43) trans-1,3-Dichloroprop...	8.182	79	13365	4.529	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.600%	
46) 2-Hexanone-d5	8.632	63	51670	49.377	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.760%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	36323	5.023	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	31042	5.383	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.600%	
Target Compounds						Qvalue
13) Acetone	2.626	43	4856	5.511	ug/L	93
22) cis-1,2-Dichloroethene	4.671	96	1131	0.169	ug/L	90
34) Trichloroethene	6.543	95	12554	1.849	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080522\
Data File : VU050101.D
Acq On : 05 Aug 2022 17:05
Operator : SY/MD
Sample : N4027-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJH6

Quant Time: Aug 06 04:36:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 06 04:32:47 2022
Response via : Initial Calibration

