

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080618\
 Data File : VU025918.D
 Acq On : 07 Aug 2018 04:38
 Operator : MD/SY
 Sample : J4344-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-0534

Quant Time: Aug 07 09:14:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 00:49:55 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	202430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	324725	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	327809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	205195	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	157029	49.79	ug/l	0.00
Spiked Amount			Recovery	=	99.58%	
35) Dibromofluoromethane	4.89	113	133060	47.11	ug/l	0.00
Spiked Amount			Recovery	=	94.22%	
50) Toluene-d8	7.57	98	478058	48.37	ug/l	0.00
Spiked Amount			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	10.31	95	192794	50.26	ug/l	0.00
Spiked Amount			Recovery	=	100.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.00	56	64112	15.00	ug/l	98
39) Methylcyclohexane	6.42	83	70702	16.26	ug/l	99
67) Ethyl Benzene	9.25	91	886672	63.33	ug/l	99
68) m/p-Xylenes	9.38	106	1380102	256.77	ug/l	99
69) o-Xylene	9.78	106	318437	60.62	ug/l	98
73) Isopropylbenzene	10.17	105	54520	3.89	ug/l	99
78) n-propylbenzene	10.59	91	48259	2.95	ug/l	100
80) 1,3,5-Trimethylbenzene	10.78	105	59909	5.04	ug/l	99
84) 1,2,4-Trimethylbenzene	11.15	105	287893	23.89	ug/l	98
85) sec-Butylbenzene	11.33	105	29814	2.07	ug/l	96
86) p-Isopropyltoluene	11.48	119	18095	1.40	ug/l	92
95) Naphthalene	13.74	128	366869	24.73	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080618\
Data File : VU025918.D
Acq On : 07 Aug 2018 04:38
Operator : MD/SY
Sample : J4344-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FL-0534

Quant Time: Aug 07 09:14:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 00:49:55 2018
Response via : Initial Calibration

