


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU080618\
Data File : VU025906.D
Acq On : 06 Aug 2018 23:33
Operator : MD/SY
Sample : BFB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U072718W.M
Title : SW846 8260
Last Update : Tue Aug 07 00:49:55 2018



AutoFind: Scans 3022, 3023, 3024; Background Corrected with Scan 3010

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	7505	PASS
75	95	30	60	53.4	20584	PASS
95	95	100	100	100.0	38528	PASS
96	95	5	9	6.9	2662	PASS
173	174	0.00	2	1.3	425	PASS
174	95	50	100	86.5	33325	PASS
175	174	5	9	7.6	2519	PASS
176	174	95	101	98.7	32893	PASS
177	176	5	9	7.3	2405	PASS