

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080718\
 Data File : VU025954.D
 Acq On : 07 Aug 2018 22:05
 Operator : MD/SY
 Sample : VU0807WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK44

Quant Time: Aug 08 08:47:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 08 06:47:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	316405	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	291769	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	122891	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	85039	48.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.00%
7) Chloroethane-d5	1.68	69	70179	53.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.74%
11) 1,1-Dichloroethene-d2	2.27	63	126060	37.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.24%
21) 2-Butanone-d5	4.18	46	136912	113.48	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.48%
24) Chloroform-d	4.65	84	175710	51.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.54%
26) 1,2-Dichloroethane-d4	5.31	65	120297	53.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.38%
32) Benzene-d6	5.35	84	341545	53.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.20%
36) 1,2-Dichloropropane-d6	6.34	67	110103	55.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.12%
41) Toluene-d8	7.57	98	312866	50.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.28%
43) trans-1,3-Dichloropropene-	7.85	79	51265	48.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.78%
47) 2-Hexanone-d5	8.31	63	90827	109.15	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.15%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	145164	52.41	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.82%
64) 1,2-Dichlorobenzene-d4	11.86	152	115047	56.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080718\
Data File : VU025954.D
Acq On : 07 Aug 2018 22:05
Operator : MD/SY
Sample : VU0807WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK44

Quant Time: Aug 08 08:47:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080718WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 08 06:47:05 2018
Response via : Initial Calibration

