

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080718\
 Data File : VU025967.D
 Acq On : 08 Aug 2018 03:22
 Operator : MD/SY
 Sample : VIBLK44
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK44

Quant Time: Aug 08 07:19:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Aug 08 06:47:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	302845	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	277172	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	120762	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	84756	50.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.00%
7) Chloroethane-d5	1.68	69	66809	53.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.16%
11) 1,1-Dichloroethene-d2	2.27	63	120789	37.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.32%
21) 2-Butanone-d5	4.18	46	122440	106.03	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.03%
24) Chloroform-d	4.65	84	169352	52.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.26%
26) 1,2-Dichloroethane-d4	5.31	65	115716	53.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.92%
32) Benzene-d6	5.35	84	331729	54.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.58%
36) 1,2-Dichloropropane-d6	6.34	67	104351	54.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.88%
41) Toluene-d8	7.57	98	300710	50.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.46%
43) trans-1,3-Dichloropropene-	7.85	79	47532	47.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.44%
47) 2-Hexanone-d5	8.31	63	82236	104.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.03%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	138390	52.59	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.18%
64) 1,2-Dichlorobenzene-d4	11.87	152	114506	57.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.80%

Target Compounds

					Ovalue
51) Chlorobenzene	9.12	112	17391	3.21 ug/L	98
63) 1,4-Dichlorobenzene	11.51	146	24944	6.78 ug/L	97
65) 1,2-Dichlorobenzene	11.88	146	19905	5.49 ug/L	99
68) 1,2,4-trichlorobenzene	13.51	180	28025	12.13 ug/L	99
70) 1,2,3-Trichlorobenzene	13.99	180	8014	3.39 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080718\
Data File : VU025967.D
Acq On : 08 Aug 2018 03:22
Operator : MD/SY
Sample : VIBLK44
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK44

Quant Time: Aug 08 07:19:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080718WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Aug 08 06:47:05 2018
Response via : Initial Calibration

