

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080718\
 Data File : VU025939.D
 Acq On : 07 Aug 2018 13:37
 Operator : MD/SY
 Sample : VU0807WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK43

Quant Time: Aug 08 04:39:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 07 12:44:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	350898	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	330558	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	171649	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	100630	51.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.50%
7) Chloroethane-d5	1.64	69	55620	38.14	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.28%
11) 1,1-Dichloroethene-d2	2.26	63	144247	38.30	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.60%
21) 2-Butanone-d5	4.18	46	143924	107.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.57%
24) Chloroform-d	4.64	84	185875	49.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.76%
26) 1,2-Dichloroethane-d4	5.31	65	127223	51.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.40%
32) Benzene-d6	5.34	84	376524	51.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.34%
36) 1,2-Dichloropropane-d6	6.33	67	118153	52.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.32%
41) Toluene-d8	7.57	98	354120	50.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.18%
43) trans-1,3-Dichloropropene-	7.85	79	61711	51.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.90%
47) 2-Hexanone-d5	8.32	63	110906	117.65	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	117.65%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	175759	56.01	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	112.02%
64) 1,2-Dichlorobenzene-d4	11.87	152	153715	54.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.42%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080718\
Data File : VU025939.D
Acq On : 07 Aug 2018 13:37
Operator : MD/SY
Sample : VU0807WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK43

Quant Time: Aug 08 04:39:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080718WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 07 12:44:32 2018
Response via : Initial Calibration

