

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080719\
 Data File : VU033615.D
 Acq On : 07 Aug 2019 14:32
 Operator : JC/SP
 Sample : K4202-09
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F2B61

Quant Time: Aug 08 03:26:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 08 03:14:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	299585	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	303966	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	136570	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	96844	50.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.16%
7) Chloroethane-d5	1.67	69	95971	44.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.94%
11) 1,1-Dichloroethene-d2	2.26	63	129887	35.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.06%
21) 2-Butanone-d5	4.16	46	167165	107.78	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.78%
24) Chloroform-d	4.62	84	185235	50.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.66%
26) 1,2-Dichloroethane-d4	5.29	65	121505	52.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.74%
32) Benzene-d6	5.32	84	354353	47.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.34%
36) 1,2-Dichloropropane-d6	6.31	67	119814	50.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.72%
41) Toluene-d8	7.55	98	299655	43.32	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.64%
43) trans-1,3-Dichloropropene-	7.83	79	53830	45.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.98%
47) 2-Hexanone-d5	8.30	63	118069	101.47	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.47%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	191677	52.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.28%
64) 1,2-Dichlorobenzene-d4	11.85	152	130904	53.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.74%


Target Compounds					Ovalue
13) Acetone	2.32	43	21059	13.631 ug/L	99




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080719\
Data File : VU033615.D
Acq On : 07 Aug 2019 14:32
Operator : JC/SP
Sample : K4202-09
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F2B61

Quant Time: Aug 08 03:26:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080519WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 08 03:14:38 2019
Response via : Initial Calibration





#13

Acetone

Concen: 13.631 ug/L

RT: 2.32 min Scan# 382

Delta R.T. -0.01 min

Lab File: VU033615.D

Acq: 07 Aug 2019 14:32

Instrument :

MSVOA_U

ClientSampleId :

F2B61

Tot Ion: 43 Resp: 21059

Ion Ratio Lower Upper

43 100

58 32.4 0.0 65.4

