

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080820\
 Data File : VU039813.D
 Acq On : 08 Aug 2020 16:05
 Operator : SY/MD
 Sample : L3591-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG0C5

Quant Time: Aug 10 05:08:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Aug 10 01:16:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	74390	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	72554	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	26386	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	22540	5.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.20%
7) Chloroethane-d5	1.92	69	19405	4.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.40%
11) 1,1-Dichloroethene-d2	2.58	63	35593	4.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.00%
20) 2-Butanone-d5	4.66	46	74322	42.07	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.14%
24) Chloroform-d	5.08	84	42799	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.40%
26) 1,2-Dichloroethane-d4	5.72	65	26438	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.40%
32) Benzene-d6	5.75	84	87037	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	6.71	67	27124	4.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.80%
41) Toluene-d8	7.91	98	73194	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	8.19	79	9551	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.64	63	50844	41.45	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.90%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	21013	3.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	24293	5.38	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.60%

Target Compounds					Ovalue
25) Chloroform	5.11	83	2469	0.252 ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU080820\
Data File : VU039813.D
Acq On : 08 Aug 2020 16:05
Operator : SY/MD
Sample : L3591-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG0C5

Quant Time: Aug 10 05:08:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Aug 10 01:16:49 2020
Response via : Initial Calibration

