

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050143.D
 Acq On : 08 Aug 2022 18:30
 Operator : SY/MD
 Sample : N4070-04DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJJ9DL

Quant Time: Aug 09 01:58:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 09 00:39:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	77196	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	79281	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	29603	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	23337	3.633	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.600%	
7) Chloroethane-d5	1.909	69	23336	4.502	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.000%	
11) 1,1-Dichloroethene-d2	2.562	65	10142	4.039	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	80.800%	
20) 2-Butanone-d5	4.636	46	98709	63.007	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	126.020%	
24) Chloroform-d	5.063	84	61472	5.144	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.800%	
26) 1,2-Dichloroethane-d4	5.706	65	33478	5.457	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.200%	
32) Benzene-d6	5.729	84	113285	4.579	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
36) 1,2-Dichloropropane-d6	6.693	67	41210	5.058	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	101.200%	
41) Toluene-d8	7.899	98	87095	3.986	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.800%	
43) trans-1,3-Dichloroprop...	8.182	79	13229	4.531	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.600%	
46) 2-Hexanone-d5	8.639	63	60188	58.135	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	116.260%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	37275	5.210	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.200%	
66) 1,2-Dichlorobenzene-d4	12.195	152	30598	5.275	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.600%	
Target Compounds						
						Qvalue
16) Methylene chloride	3.044	84	8456	1.037	ug/L	92
22) cis-1,2-Dichloroethene	4.668	96	6486	0.978	ug/L	91
34) Trichloroethene	6.542	95	41846	6.229	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050143.D
Acq On : 08 Aug 2022 18:30
Operator : SY/MD
Sample : N4070-04DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJJ9DL

Quant Time: Aug 09 01:58:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 09 00:39:33 2022
Response via : Initial Calibration

