

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050150.D
 Acq On : 08 Aug 2022 21:48
 Operator : SY/MD
 Sample : VU0808WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK103

Quant Time: Aug 09 04:06:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 09 03:59:45 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.250	114	91303	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.420	117	89076	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.816	152	34284	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.597	65	42610	5.608	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.200%	
7) Chloroethane-d5		1.909	69	35096	5.725	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.600%	
11) 1,1-Dichloroethene-d2		2.565	65	16260	5.475	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	109.600%	
20) 2-Butanone-d5		4.633	46	107099	57.800	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.600%	
24) Chloroform-d		5.067	84	76106	5.385	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%	
26) 1,2-Dichloroethane-d4		5.703	65	38926	5.364	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%	
32) Benzene-d6		5.729	84	154097	5.544	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.800%	
36) 1,2-Dichloropropane-d6		6.694	67	50783	5.548	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.000%	
41) Toluene-d8		7.899	98	123242	5.020	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%	
43) trans-1,3-Dichloroprop...		8.182	79	15794	4.815	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.200%	
46) 2-Hexanone-d5		8.636	63	66814	57.438	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.880%	
56) 1,1,2,2-Tetrachloroeth...		10.758	84	40651	5.057	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.200%	
66) 1,2-Dichlorobenzene-d4		12.195	152	38461	5.726	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.600%	
Target Compounds							
16) Methylene chloride		3.044	84	6272	0.650	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050150.D
Acq On : 08 Aug 2022 21:48
Operator : SY/MD
Sample : VU0808WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK103

Quant Time: Aug 09 04:06:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 09 03:59:45 2022
Response via : Initial Calibration

