

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050159.D
 Acq On : 09 Aug 2022 01:55
 Operator : SY/MD
 Sample : N4075-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUV4

Quant Time: Aug 09 05:18:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 09 03:59:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	34749	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	33563	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	11515	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	12818	4.433	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.600%	
7) Chloroethane-d5	1.909	69	12017	5.151	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	103.000%	
11) 1,1-Dichloroethene-d2	2.565	65	5470	4.840	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	96.800%	
20) 2-Butanone-d5	4.623	46	73760	104.594	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	209.180%#	
24) Chloroform-d	5.063	84	29799	5.540	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.800%	
26) 1,2-Dichloroethane-d4	5.703	65	18753	6.790	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	135.800%#	
32) Benzene-d6	5.729	84	53350	5.094	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.800%	
36) 1,2-Dichloropropane-d6	6.690	67	19729	5.720	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	114.400%	
41) Toluene-d8	7.899	98	39046	4.221	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.400%	
43) trans-1,3-Dichloroprop...	8.179	79	6076	4.916	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	98.400%	
46) 2-Hexanone-d5	8.636	63	39838	90.893	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	181.780%#	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	23668	7.814	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	156.200%#	
66) 1,2-Dichlorobenzene-d4	12.195	152	12905	5.720	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	114.400%	
Target Compounds						
					Qvalue	
13) Acetone	2.626	43	3967	10.101	ug/L	98
14) Carbon disulfide	2.790	76	4026	0.469	ug/L	98
16) Methylene chloride	3.044	84	1567	0.427	ug/L	94
18) trans-1,2-Dichloroethene	3.356	96	3642	1.354	ug/L	95
22) cis-1,2-Dichloroethene	4.668	96	8989	3.011	ug/L	96
34) Trichloroethene	6.546	95	1081	0.380	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050159.D
Acq On : 09 Aug 2022 01:55
Operator : SY/MD
Sample : N4075-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUV4

Quant Time: Aug 09 05:18:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 09 03:59:45 2022
Response via : Initial Calibration

