

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050160.D
 Acq On : 09 Aug 2022 02:19
 Operator : SY/MD
 Sample : N4075-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUV5

Quant Time: Aug 09 05:18:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 09 03:59:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	46487	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	44787	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	16102	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	16375	4.233	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.600%
7) Chloroethane-d5	1.909	69	14784	4.737	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.800%
11) 1,1-Dichloroethene-d2	2.562	65	6508	4.304	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.000%
20) 2-Butanone-d5	4.623	46	87668	92.926	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	185.860%#
24) Chloroform-d	5.063	84	38638	5.369	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.400%
26) 1,2-Dichloroethane-d4	5.703	65	23226	6.286	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.800%
32) Benzene-d6	5.729	84	69048	4.940	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
36) 1,2-Dichloropropane-d6	6.690	67	25861	5.619	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.400%
41) Toluene-d8	7.899	98	50629	4.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
43) trans-1,3-Dichloroprop...	8.182	79	7841	4.754	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.000%
46) 2-Hexanone-d5	8.632	63	49492	84.621	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	169.240%#
56) 1,1,2,2-Tetrachloroeth...	10.758	84	30079	7.442	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	148.800%#
66) 1,2-Dichlorobenzene-d4	12.195	152	17898	5.673	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.400%
Target Compounds						
					Qvalue	
13) Acetone	2.629	43	2502	4.762	ug/L	93
14) Carbon disulfide	2.790	76	3426	0.299	ug/L	97
16) Methylene chloride	3.044	84	1105	0.225	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050160.D
Acq On : 09 Aug 2022 02:19
Operator : SY/MD
Sample : N4075-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUV5

Quant Time: Aug 09 05:18:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 09 03:59:45 2022
Response via : Initial Calibration

