

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050198.D
 Acq On : 09 Aug 2022 23:09
 Operator : SY/MD
 Sample : N4076-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBUU0

Quant Time: Aug 10 06:34:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 10 05:49:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	60951	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	58201	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	22420	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	39850	7.856	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	157.200%#	
7) Chloroethane-d5	1.906	69	28186	6.887	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	137.800%#	
11) 1,1-Dichloroethene-d2	2.562	65	13276	6.697	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	134.000%#	
20) 2-Butanone-d5	4.623	46	104786	84.713	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	169.420%#	
24) Chloroform-d	5.063	84	56115	5.948	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	119.000%	
26) 1,2-Dichloroethane-d4	5.703	65	31945	6.594	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	131.800%#	
32) Benzene-d6	5.729	84	111750	6.153	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	123.000%	
36) 1,2-Dichloropropane-d6	6.694	67	37468	6.265	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	125.200%	
41) Toluene-d8	7.899	98	87603	5.461	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.200%	
43) trans-1,3-Dichloroprop...	8.179	79	12677	5.915	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	118.200%	
46) 2-Hexanone-d5	8.632	63	65752	86.512	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	173.020%#	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	36574	6.963	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	139.200%#	
66) 1,2-Dichlorobenzene-d4	12.195	152	28890	6.577	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	131.600%#	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.578	96	15483	3.675	ug/L	# 83
13) Acetone	2.626	43	5542	8.045	ug/L	85
14) Carbon disulfide	2.790	76	9400	0.625	ug/L	98
18) trans-1,2-Dichloroethene	3.353	96	214821	45.539	ug/L	95
19) 1,1-Dichloroethane	3.871	63	32357	3.622	ug/L	99
22) cis-1,2-Dichloroethene	4.665	96	1540568	294.233	ug/L	100
33) Benzene	5.774	78	8735	0.428	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050198.D
Acq On : 09 Aug 2022 23:09
Operator : SY/MD
Sample : N4076-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBUU0

Quant Time: Aug 10 06:34:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 10 05:49:54 2022
Response via : Initial Calibration

