

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050157.D
 Acq On : 09 Aug 2022 01:06
 Operator : SY/MD
 Sample : N4070-22
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJM1

Quant Time: Aug 09 05:18:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 09 03:59:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	75897	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	77865	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	29061	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	21474	3.400	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	68.000%	
7) Chloroethane-d5	1.909	69	20698	4.062	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	81.200%	
11) 1,1-Dichloroethene-d2	2.565	65	8329	3.374	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.400%	
20) 2-Butanone-d5	4.623	46	87942	57.095	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	114.200%	
24) Chloroform-d	5.063	84	55656	4.737	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.800%	
26) 1,2-Dichloroethane-d4	5.703	65	29942	4.964	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.200%	
32) Benzene-d6	5.729	84	100937	4.154	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.000%	
36) 1,2-Dichloropropane-d6	6.694	67	37805	4.725	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.400%	
41) Toluene-d8	7.899	98	76272	3.554	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.000%	
43) trans-1,3-Dichloroprop...	8.182	79	11157	3.891	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.800%	
46) 2-Hexanone-d5	8.632	63	50987	50.143	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	100.280%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	35261	5.018	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.400%	
66) 1,2-Dichlorobenzene-d4	12.195	152	28447	4.996	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.000%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.575	96	984	0.188	ug/L #	1
13) Acetone	2.626	43	5172	6.029	ug/L	100
19) 1,1-Dichloroethane	3.877	63	1023	0.092	ug/L #	87
22) cis-1,2-Dichloroethene	4.665	96	21965	3.369	ug/L	97
34) Trichloroethene	6.542	95	71926	10.901	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050157.D
Acq On : 09 Aug 2022 01:06
Operator : SY/MD
Sample : N4070-22
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJM1

Quant Time: Aug 09 05:18:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 09 03:59:45 2022
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 0.188 ug/L

RT: 2.575 min Scan# 384

Delta R.T. -0.003 min

Lab File: VU050157.D

Acq: 09 Aug 2022 01:06

Instrument :





MSVOA_U

ClientSampleId :

GBJM1

Tgt Ion: 96 Resp: 984

Ion	Ratio	Lower	Upper
96	100		
61	311.1	122.7	227.9#
63	1850.2	85.1	158.1#



#13

Acetone

Concen: 6.029 ug/L

RT: 2.626 min Scan# 400




Delta R.T. -0.000 min


Lab File: VU050157.D

Acq: 09 Aug 2022 01:06

Tgt Ion: 43 Resp: 5172

Ion	Ratio	Lower	Upper
43	100		
58	32.1	0.0	64.0





#19

1,1-Dichloroethane

Concen: 0.092 ug/L

RT: 3.877 min Scan# 71

Delta R.T. 0.009 min

Lab File: VU050157.D

Acq: 09 Aug 2022 01:06

Instrument :

MSVOA_U

ClientSampleId :

GBJM1





Tgt Ion: 63 Resp: 1023

Ion Ratio Lower Upper

63 100

65 34.8 22.7 42.3

83 0.0 9.8 18.2#



#22

cis-1,2-Dichloroethene

Concen: 3.369 ug/L

RT: 4.665 min Scan# 1034

Delta R.T. -0.003 min

Lab File: VU050157.D

Acq: 09 Aug 2022 01:06




Tgt Ion: 96 Resp: 21965


Ion Ratio Lower Upper

96 100

61 122.6 89.0 165.4

98 64.7 45.2 84.0





#34
 Trichloroethene
 Concen: 10.901 ug/L
 RT: 6.542 min Scan# 1
 Delta R.T. -0.000 min
 Lab File: VU050157.D
 Acq: 09 Aug 2022 01:06

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJM1

Tgt Ion: 95 Resp: 71926

Ion	Ratio	Lower	Upper
95	100		
97	65.7	45.6	84.8
132	89.9	65.0	120.6
130	96.3	69.3	128.7

