

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
 Data File : VU050175.D
 Acq On : 09 Aug 2022 11:38
 Operator : SY/MD
 Sample : N4070-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBJL4

Quant Time: Aug 10 04:46:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Aug 09 03:59:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	83066	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	83405	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	31835	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	21121	3.055	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.200%
7) Chloroethane-d5	1.909	69	21666	3.885	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.600%
11) 1,1-Dichloroethene-d2	2.562	65	8888	3.290	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.800%
20) 2-Butanone-d5	4.626	46	98610	58.496	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.000%
24) Chloroform-d	5.063	84	58655	4.562	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.200%
26) 1,2-Dichloroethane-d4	5.703	65	32485	4.921	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	5.729	84	105999	4.073	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.400%
36) 1,2-Dichloropropane-d6	6.694	67	40073	4.675	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.600%
41) Toluene-d8	7.899	98	82126	3.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.400%
43) trans-1,3-Dichloroprop...	8.182	79	13073	4.256	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.200%
46) 2-Hexanone-d5	8.636	63	62890	57.741	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.480%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	37512	4.984	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	12.195	152	29495	4.729	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080822\
Data File : VU050175.D
Acq On : 09 Aug 2022 11:38
Operator : SY/MD
Sample : N4070-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBJL4

Quant Time: Aug 10 04:46:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Aug 09 03:59:45 2022
Response via : Initial Calibration

