

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080824\
 Data File : VU060365.D
 Acq On : 08 Aug 2024 18:52
 Operator : MD/SY
 Sample : P3401-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1GG4

Quant Time: Aug 09 00:05:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Aug 08 23:46:27 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.245	114	105730	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.412	117	114487	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.807	152	52468	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.595	65	41705	5.025	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.600%	
7) Chloroethane-d5		1.911	69	39058	5.219	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.400%	
11) 1,1-Dichloroethene-d2		2.563	65	16820	4.762	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.200%	
20) 2-Butanone-d5		4.634	46	88666	50.341	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.680%	
24) Chloroform-d		5.055	84	84810	5.161	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%	
26) 1,2-Dichloroethane-d4		5.698	65	38897	5.104	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.000%	
32) Benzene-d6		5.721	84	165214	4.785	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6		6.685	67	55723	5.203	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.000%	
41) Toluene-d8		7.894	98	140428	4.556	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.200%	
43) trans-1,3-Dichloroprop...		8.177	79	15640	4.517	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.400%	
46) 2-Hexanone-d5		8.631	63	71002	46.210	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.420%	
56) 1,1,2,2-Tetrachloroeth...		10.749	84	43192	4.915	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.400%	
66) 1,2-Dichlorobenzene-d4		12.187	152	54760	5.666	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.400%	
Target Compounds							
13) Acetone		2.650	43	30106	25.031	ug/L	98
16) Methylene chloride		3.039	84	2077	0.210	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080824\
Data File : VU060365.D
Acq On : 08 Aug 2024 18:52
Operator : MD/SY
Sample : P3401-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1GG4

Quant Time: Aug 09 00:05:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Aug 08 23:46:27 2024
Response via : Initial Calibration

