

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080918\
 Data File : VU025995.D
 Acq On : 09 Aug 2018 14:46
 Operator : MD/SY
 Sample : J4361-10 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4ZB7

Quant Time: Aug 10 06:03:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 10 05:14:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	300338	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	286084	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	144329	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	96796	44.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.18%
7) Chloroethane-d5	1.68	69	80749	46.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.78%
11) 1,1-Dichloroethene-d2	2.27	63	153128	36.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.42%
21) 2-Butanone-d5	4.18	46	177335	96.92	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.92%
24) Chloroform-d	4.65	84	219187	47.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.16%
26) 1,2-Dichloroethane-d4	5.31	65	151717	48.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.48%
32) Benzene-d6	5.34	84	427389	48.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.22%
36) 1,2-Dichloropropane-d6	6.33	67	139731	48.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.22%
41) Toluene-d8	7.57	98	398693	46.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.82%
43) trans-1,3-Dichloropropene-	7.85	79	70839	48.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.82%
47) 2-Hexanone-d5	8.31	63	124791	95.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.58%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	198414	47.20	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	175928	49.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.90%

Target Compounds

						Ovalue
13) Acetone	2.32	43	6925	6.09	ug/L	95
42) Toluene	7.64	91	33601	3.89	ug/L	95
51) Chlorobenzene	9.12	112	816678	140.04	ug/L	99
62) 1,3-Dichlorobenzene	11.42	146	107639	23.96	ug/L	99
63) 1,4-Dichlorobenzene	11.51	146	904913	192.12	ug/L	99
65) 1,2-Dichlorobenzene	11.88	146	1244001	260.10	ug/L	99
68) 1,2,4-trichlorobenzene	13.51	180	1880817	662.30	ug/L	99
70) 1,2,3-Trichlorobenzene	13.99	180	451894	148.05	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080918\
Data File : VU025995.D
Acq On : 09 Aug 2018 14:46
Operator : MD/SY
Sample : J4361-10 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4ZB7

Quant Time: Aug 10 06:03:22 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080918WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 10 05:14:50 2018
Response via : Initial Calibration

