

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080918\
 Data File : VU026026.D
 Acq On : 10 Aug 2018 03:44
 Operator : MD/SY
 Sample : J4401-11 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A4ZF1

Quant Time: Aug 10 08:26:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080918WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Aug 10 07:39:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	295846	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	279432	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	135410	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	77014	35.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.22%
7) Chloroethane-d5	1.68	69	68020	39.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.34%
11) 1,1-Dichloroethene-d2	2.27	63	117250	28.53	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.06%#
21) 2-Butanone-d5	4.18	46	151243	83.91	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.91%
24) Chloroform-d	4.65	84	190954	41.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.28%
26) 1,2-Dichloroethane-d4	5.31	65	132912	42.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.82%
32) Benzene-d6	5.35	84	362639	41.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.58%
36) 1,2-Dichloropropane-d6	6.33	67	120172	42.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.60%
41) Toluene-d8	7.57	98	324609	39.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.20%#
43) trans-1,3-Dichloropropene-	7.85	79	55567	39.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.56%
47) 2-Hexanone-d5	8.31	63	106973	83.89	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.89%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	170327	41.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.96%
64) 1,2-Dichlorobenzene-d4	11.86	152	143031	42.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.70%

Target Compounds


						Ovalue
13) Acetone	2.32	43	41866	37.39	ug/L	100
22) 2-Butanone	4.27	43	6986	4.31	ug/L	# 65
51) Chlorobenzene	9.12	112	453875	79.68	ug/L	100
62) 1,3-Dichlorobenzene	11.42	146	76902	18.24	ug/L	98
63) 1,4-Dichlorobenzene	11.51	146	400418	90.61	ug/L	98
65) 1,2-Dichlorobenzene	11.88	146	448567	99.97	ug/L	99
68) 1,2,4-trichlorobenzene	13.51	180	686480	257.66	ug/L	99
70) 1,2,3-Trichlorobenzene	13.99	180	150305	52.49	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU080918\
Data File : VU026026.D
Acq On : 10 Aug 2018 03:44
Operator : MD/SY
Sample : J4401-11 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A4ZF1

Quant Time: Aug 10 08:26:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080918WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Aug 10 07:39:13 2018
Response via : Initial Calibration





#13

Acetone

Concen: 37.39 ug/L

RT: 2.32 min Scan# 538

Delta R.T. 0.00 min

Lab File: VU026026.D

Acq: 10 Aug 2018 03:44

Instrument :

MSVOA_U

ClientSampleId :





A4ZF1

Tot Ion: 43 Resp: 41866

Ion Ratio Lower Upper

43 100

58 30.8 0.0 61.8



#22

2-Butanone

Concen: 4.31 ug/L

RT: 4.27 min Scan# 1144

Delta R.T. 0.01 min

Lab File: VU026026.D




Acq: 10 Aug 2018 03:44


Tgt Ion: 43 Resp: 6986

Ion Ratio Lower Upper

43 100

72 9.0 13.5 40.5#





#51

Chlorobenzene

Concen: 79.68 ug/L

RT: 9.12 min Scan# 2653

Delta R.T. -0.00 min

Lab File: VU026026.D

Acq: 10 Aug 2018 03:44

Instrument :

MSVOA_U

ClientSampleId :

A4ZF1


Tot Ion:112 Resp: 453875

Ion Ratio Lower Upper

112 100

114 31.8 22.7 42.3




77 56.5 45.2 67.8



Abundance Ion 112.00 (111.70 to 112.70): V

Ion 114.00 (113.70 to 114.70): V

Ion 77.00 (76.70 to 77.70): VU0



#62

1,3-Dichlorobenzene

Concen: 18.24 ug/L

RT: 11.42 min Scan# 3368

Delta R.T. 0.00 min

Lab File: VU026026.D

Acq: 10 Aug 2018 03:44


Tgt Ion:146 Resp: 76902

Ion Ratio Lower Upper

146 100

111 41.1 27.9 51.9



148 66.7 46.0 85.4




Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V









#63
1,4-Dichlorobenzene
Concen: 90.61 ug/L
RT: 11.51 min Scan# 3396
Delta R.T. 0.00 min
Lab File: VU026026.D
Acq: 10 Aug 2018 03:44




Instrument :
MSVOA_U
ClientSampleId :
A4ZF1


Tot Ion	Ratio	Lower	Upper
146	100		
111	37.8	26.4	49.0
148	65.0	43.8	81.4



#65
1,2-Dichlorobenzene
Concen: 99.97 ug/L
RT: 11.88 min Scan# 3512
Delta R.T. 0.00 min
Lab File: VU026026.D
Acq: 10 Aug 2018 03:44

Tgt Ion	Ratio	Lower	Upper
146	100		
111	41.0	32.4	48.6
148	63.1	50.1	75.1





#68

1,2,4-trichlorobenzene

Concen: 257.66 ug/L

RT: 13.51 min Scan# 4017

Delta R.T. 0.00 min

Lab File: VU026026.D

Acq: 10 Aug 2018 03:44

Instrument :

MSVOA_U

ClientSampleId :

A4ZF1


Tot Ion:180 Resp: 686480

Ion Ratio Lower Upper

180 100

182 94.7 76.5 114.7




145 29.9 24.2 36.4



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V



#70

1,2,3-Trichlorobenzene

Concen: 52.49 ug/L

RT: 13.99 min Scan# 4168

Delta R.T. 0.00 min

Lab File: VU026026.D

Acq: 10 Aug 2018 03:44


Tgt Ion:180 Resp: 150305

Ion Ratio Lower Upper

180 100

182 96.2 76.5 114.7

145 31.8 24.7 37.1



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V

