

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080924\
 Data File : VU060390.D
 Acq On : 09 Aug 2024 22:59
 Operator : MD/SY
 Sample : P3540-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B45

Quant Time: Aug 10 06:01:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Aug 10 05:58:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	97867	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	103630	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	40695	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	30572	3.980	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.600%
7) Chloroethane-d5	1.910	69	31833	4.595	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.560	65	13662	4.179	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.600%
20) 2-Butanone-d5	4.637	46	86052	52.782	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.560%
24) Chloroform-d	5.055	84	72435	4.762	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
26) 1,2-Dichloroethane-d4	5.695	65	35238	4.996	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.000%
32) Benzene-d6	5.721	84	134712	4.310	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
36) 1,2-Dichloropropane-d6	6.682	67	46750	4.822	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.400%
41) Toluene-d8	7.894	98	104072	3.730	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.600%
43) trans-1,3-Dichloroprop...	8.177	79	12840	4.097	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.630	63	64483	46.364	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.720%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	38192	4.801	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	39533	5.273	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU080924\
Data File : VU060390.D
Acq On : 09 Aug 2024 22:59
Operator : MD/SY
Sample : P3540-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B45

Quant Time: Aug 10 06:01:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR080824WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Aug 10 05:58:50 2024
Response via : Initial Calibration

