

Data File : VU033732.D
Acq On : 10 Aug 2019 22:03
Operator : JC/SP
Sample : K4289-09RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F2B29RE

Quant Time: Aug 12 01:48:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
Quant Title : VOC Analysis
QLast Update : Sun Aug 11 04:51:09 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	292311	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	291778	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	142178	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	37006	17.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	34.40%#
7) Chloroethane-d5	1.67	69	41101	19.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	38.12%#
11) 1,1-Dichloroethene-d2	2.26	63	51505	14.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	28.48%#
21) 2-Butanone-d5	4.15	46	157505	118.07	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.07%
24) Chloroform-d	4.62	84	118907	37.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.38%
26) 1,2-Dichloroethane-d4	5.29	65	79723	39.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.76%
32) Benzene-d6	5.32	84	192970	28.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	56.04%#
36) 1,2-Dichloropropane-d6	6.31	67	78996	37.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	74.76%
41) Toluene-d8	7.55	98	151866	23.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	46.90%#
43) trans-1,3-Dichloropropene-	7.83	79	33296	31.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.40%
47) 2-Hexanone-d5	8.29	63	110806	101.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.23%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	155932	49.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.76%
64) 1,2-Dichlorobenzene-d4	11.84	152	94935	39.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.34%#

Target Compounds

						Qvalue
5) Vinyl chloride	1.39	62	61863	25.140	ug/L	91
17) trans-1,2-Dichloroethene	2.97	96	5492	2.981	ug/L	90
20) cis-1,2-Dichloroethene	4.20	96	194415	92.519	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU033732.D
Acq On : 10 Aug 2019 22:03
Operator : JC/SP
Sample : K4289-09RE
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F2B29RE

Quant Time: Aug 12 01:48:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM080919WMA.M
Quant Title : VOC Analysis
QLast Update : Sun Aug 11 04:51:09 2019
Response via : Initial Calibration

