

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081020\
 Data File : VU039825.D
 Acq On : 10 Aug 2020 15:35
 Operator : SY/MD
 Sample : L3572-16
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Aug 11 05:17:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 11 04:57:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	86123	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	83185	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	40729	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	20638	4.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.80%
7) Chloroethane-d5	1.92	69	22031	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.60%
11) 1,1-Dichloroethene-d2	2.58	63	30713	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	4.65	46	68182	33.33	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	66.66%
24) Chloroform-d	5.08	84	41705	3.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.80%
26) 1,2-Dichloroethane-d4	5.72	65	25296	3.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.20%
32) Benzene-d6	5.74	84	83340	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.20%
36) 1,2-Dichloropropane-d6	6.70	67	27525	4.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.20%
41) Toluene-d8	7.90	98	68885	3.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.40%
43) trans-1,3-Dichloropropene-	8.19	79	8881	3.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.40%
46) 2-Hexanone-d5	8.64	63	46818	33.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.58%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	21411	3.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	65.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	29941	4.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081020\
Data File : VU039825.D
Acq On : 10 Aug 2020 15:35
Operator : SY/MD
Sample : L3572-16
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Aug 11 05:17:34 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 11 04:57:12 2020
Response via : Initial Calibration

