

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU081020\
 Data File : VU039834.D
 Acq On : 10 Aug 2020 19:23
 Operator : SY/MD
 Sample : L3578-04
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Aug 11 06:16:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Aug 11 04:57:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	75316	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	75904	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	36223	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	21672	5.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.60%
7) Chloroethane-d5	1.92	69	20794	4.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.60%
11) 1,1-Dichloroethene-d2	2.58	63	31720	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	4.66	46	80138	44.80	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.60%
24) Chloroform-d	5.08	84	39309	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.40%
26) 1,2-Dichloroethane-d4	5.72	65	26391	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.20%
32) Benzene-d6	5.75	84	82448	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	27380	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.91	98	68492	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloropropene-	8.19	79	9238	3.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.00%
46) 2-Hexanone-d5	8.64	63	56048	43.68	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.36%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	24362	4.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	30484	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%


Target Compounds					Ovalue
25) Chloroform	5.11	83	4935	0.497 ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU081020\
Data File : VU039834.D
Acq On : 10 Aug 2020 19:23
Operator : SY/MD
Sample : L3578-04
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Aug 11 06:16:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR080320WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 11 04:57:12 2020
Response via : Initial Calibration





#25
 Chloroform
 Concen: 0.497 ug/L
 RT: 5.11 min Scan# 1172
 Delta R.T. 0.00 min
 Lab File: VU039834.D
 Acq: 10 Aug 2020 19:23

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Tot Ion: 83 Resp: 4935
 Ion Ratio Lower Upper
 83 100
 85 64.2 42.1 78.3

