

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081021\
 Data File : VU044409.D
 Acq On : 10 Aug 2021 13:36
 Operator : SY/MD
 Sample : M3344-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AS5

Quant Time: Aug 11 01:09:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Aug 11 01:08:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	163135	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	162497	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	75539	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	47546	3.488	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.800%
7) Chloroethane-d5	1.912	69	59199	5.059	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.200%
11) 1,1-Dichloroethene-d2	2.568	65	22941	3.996	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.000%
20) 2-Butanone-d5	4.629	46	161059	39.848	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.700%
24) Chloroform-d	5.070	84	112174	4.729	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.710	65	64273	4.854	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
32) Benzene-d6	5.735	84	236549	4.646	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.697	67	73939	4.673	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.902	98	212096	4.654	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
43) trans-1,3-Dichloroprop...	8.185	79	22030	3.576	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.600%
46) 2-Hexanone-d5	8.639	63	142122	38.216	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.440%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	62988	4.452	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.000%
66) 1,2-Dichlorobenzene-d4	12.198	152	80206	5.514	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.200%
Target Compounds						
					Qvalue	
13) Acetone	2.633	43	5963	3.098	ug/L	99
25) Chloroform	5.096	83	3744	0.176	ug/L	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081021\
Data File : VU044409.D
Acq On : 10 Aug 2021 13:36
Operator : SY/MD
Sample : M3344-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AS5

Quant Time: Aug 11 01:09:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR072321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Aug 11 01:08:01 2021
Response via : Initial Calibration

