

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU081021\  
 Data File : VU044411.D  
 Acq On : 10 Aug 2021 14:33  
 Operator : SY/MD  
 Sample : M3344-03  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampled :  
 C0AS6

Quant Time: Aug 11 01:09:57 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR072321WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Aug 11 01:08:01 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	158961	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	158972	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	76595	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	47969	3.612	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.200%	
7) Chloroethane-d5	1.912	69	59259	5.197	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	104.000%	
11) 1,1-Dichloroethene-d2	2.568	65	23267	4.159	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	83.200%	
20) 2-Butanone-d5	4.629	46	162301	41.209	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	82.420%	
24) Chloroform-d	5.070	84	109770	4.749	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.000%	
26) 1,2-Dichloroethane-d4	5.710	65	62721	4.861	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.200%	
32) Benzene-d6	5.735	84	241716	4.852	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
36) 1,2-Dichloropropane-d6	6.700	67	76144	4.919	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.400%	
41) Toluene-d8	7.906	98	213377	4.786	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.800%	
43) trans-1,3-Dichloroprop...	8.185	79	22796	3.783	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.600%	
46) 2-Hexanone-d5	8.639	63	145189	39.906	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	79.820%	
56) 1,1,2,2-Tetrachloroeth...	10.764	84	62144	4.490	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.800%	
66) 1,2-Dichlorobenzene-d4	12.201	152	80869	5.483	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	109.600%	
-----						
Target Compounds						Qvalue
13) Acetone	2.632	43	4650	2.479	ug/L	91
14) Carbon disulfide	2.796	76	26193	0.768	ug/L	99
25) Chloroform	5.095	83	9095	0.439	ug/L	93
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU081021\  
 Data File : VU044411.D  
 Acq On : 10 Aug 2021 14:33  
 Operator : SY/MD  
 Sample : M3344-03  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampled :  
 C0AS6

Quant Time: Aug 11 01:09:57 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR072321WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Aug 11 01:08:01 2021  
 Response via : Initial Calibration

