

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081022\
 Data File : VU050221.D
 Acq On : 10 Aug 2022 16:04
 Operator : SY/MD
 Sample : N4043-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PW-B6-L10-M-080322

Quant Time: Aug 10 16:28:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081022DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Aug 10 15:54:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.115	96	25735	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.636	95	6490	0.735	ug/l	0.00
Spiked Amount 1.000			Recovery	=	74.000%	
68) 1,2-Dichlorobenzene-d4	12.195	152	7641	0.856	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.000%	
Target Compounds						
						Qvalue
3) Chloromethane	1.517	50	1950	0.168	ug/l	96
13) Acetone	2.658	43	2444	1.807	ug/l	97
14) Carbon Disulfide	2.784	76	1671	0.065	ug/l	# 77
15) Methylene Chloride	3.041	84	3233	0.316	ug/l	90
69) Isopropylbenzene	10.488	105	619	0.022	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU081022\
Data File : VU050221.D
Acq On : 10 Aug 2022 16:04
Operator : SY/MD
Sample : N4043-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PW-B6-L10-M-080322

Quant Time: Aug 10 16:28:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081022DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Aug 10 15:54:10 2022
Response via : Initial Calibration

